

## **Analytical Test Report**

Science First. ORELAP ID: 4101 OLCC#: 010-10233216933

20756 High Desert Court STE 6, Bend, OR 97701 (541) 299-2008

### Strawberry Mojito Kegs FD

**Lab ID:** B240437-02

**Ablis CBD** 

Harvest / Process Lot: 04/18/24 METRC Batch ID: KEG.SMJ.027 METRC Sample Tag: KEG.SMJ.027

**Date Sampled:** 04/19/24 **Date Printed:** 04/30/24

#### **Potency Analysis**

Analytical Method: JA-Potency-Proprietary BBLB SOP 009 - Potency Analysis by HPLC-UV-Vis

Cannabinoids	mg/g	LOQ Notes
THCA	< LOQ	0.002646
delta 9-THC	< LOQ	0.002646
delta 8-THC	< LOQ	0.002646
CBDA	< LOQ	0.002646
CBGA	< LOQ	0.002646
CBD	0.1550	0.002646
CBG	< LOQ	0.002646
CBN	< LOQ	0.002646
CBC	< LOQ	0.002646
THCV	< LOQ	0.002646
Total CBG	< LOQ	0.001323
Total Cannabinoids	0.1550	0.001323


Total THC < LOQ mg/g

Total CBD 0.1550 mg/g

<LOQ - Results below the Limit of Quantitation

ORELAP accredited cannabinoid analytes include only CBDA, CBD, THCA, delta-9-THC, and delta-8-THC. Acid form of THC/CBD are decarboxylated by heat, lose 12% of original mass as CO2. Result = \*bioactive\*

<sup>&</sup>quot;Total" Cannabinoid accounts for decarboxylation and moisture content. Total THC = [(THCA×0.877) +  $\Delta$ 9THC] / (100%-MC)



## **Analytical Test Report**

Science First. ORELAP ID: 4101 OLCC#: 010-10233216933

20756 High Desert Court STE 6, Bend, OR 97701 (541) 299-2008

#### **Strawberry Mojito Kegs FD**

Ablis CBD

Laboratory ID: B240437-02


# Quality Control Potency

Batch: B24D224 - BB Potency

Blank(B24D224-BLK1)							
Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
delta 9-THC	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
CBGA	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
CBDA	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
CBD	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
CBN	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
CBG	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
delta 8-THC	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
CBC	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
Total CBG	< LOQ	0.01992	mg/g		04/23/24 13:27	04/24/24 08:44	
THCV	< LOQ	0.03983	mg/g		04/23/24 13:27	04/24/24 08:44	
LCS(B24D224-I	BS1)						
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes

LCS(B24D224-	BS1)						
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	100	0.03862	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
delta 9-THC	99.0	0.03862	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
CBDA	102	0.03862	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
CBD	102	0.03862	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
delta 8-THC	95.4	0.03862	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	

LCS Dup(B24D224-BSD1)							
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	101	0.03996	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
delta 9-THC	99.3	0.03996	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
CBDA	103	0.03996	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
CBD	103	0.03996	mg/g	90-110	04/23/24 13:27	04/24/24 08:44	
delta 8-THC	97 4	0 03996	ma/a	90-110	04/23/24 13:27	04/24/24 08:44	



## **Analytical Test Report**

Science First. ORELAP ID: 4101 OLCC#: 010-10233216933

20756 High Desert Court STE 6, Bend, OR 97701 (541) 299-2008

#### Strawberry Mojito Kegs FD

Ablis CBD

Laboratory ID: B240437-02

#### **Notes and Definitions**

- B Analyte detected in method blank, but not associated samples.
- B2 Analyte detected in sample and associate methold blank.
- C Interference due to co-elution.
- D Initial result exceeded calibration range, reported data are based on analysis of a dilution.
- H Non-homogenous samle matrix affecting RPD and/or QC.
- I Manual Integration was performed.
- L Duplicate sample relative precent difference (RPD) exceeds QC limits.
- M Anomalous results due to matrix interference
- Peaks manually split.
- Q1 QC out of limits but still oK
- Q2 Quality Control outside QC limits. Data considered estimate.
- Q3 CCV was above the acceptance criteria. Non-detect samples are considered acceptable.
- Q4 CCV was below the acceptance criteria, however the sample still exceeds the regulatory limit.
- R Marginal Exceedence.
- U Reported result is an estimate. The analyte was detected above the calibration range.
- X Problems with initial analysis, reported data are from reinjection of prepared sample.
- <LOQ Results below the Limit of Quantitation Compound not detected